

**TEST FACILITY**

---

**NAMSA**  
6750 Wales Road  
Northwood, OH 43619  
419.666.9455

**CLIENT**

---

**SMOOTH-ON, INC**  
2000 Saint John Street  
Easton, PA 18042

---

<b>Test Report No: 14T_56434_03</b>	<b>Date: December 29, 2014</b>
-------------------------------------	--------------------------------

**SAMPLE ID:** The client identified the following test material as “**Dragon Skin FX Pro**”

**SAMPLING DETAIL:** Test samples were submitted to the laboratory directly by the client. No special sampling conditions or sample preparation were observed by NAMSA.

**DATE OF RECEIPT:** Samples were received at NAMSA facilities on October 21, 2014.

**TESTING PERIOD:** November 10, 2014

**AUTHORIZATION:** Signed work order 116924.0 signed by Rodney C. Conn

**TEST REQUESTED:** Evaluate for the potential to elicit delayed dermal contact sensitization. This study was conducted based on the requirements of ISO 10993-10, Biological evaluation of medical devices, Part 10: Tests for irritation and skin sensitization.

**TEST RESULTS:** The test article showed no evidence of causing delayed dermal contact sensitization.

**Prepared For:**

**Rodney C. Conn**  
**Technical Compliance Specialist**  
**SMOOTH-ON, INC.**